

To whom it may concern

May 2024
Sugisaki Meter Co., Ltd.

DI-12 series Calibration accuracy change notice

I would like to congratulate you on your continued prosperity. Thank you very much for your continued patronage.

Due to various circumstances, we have decided to change the calibration accuracy of the "Slip Torque Tester DI-12 Series".

We apologize for any inconvenience this may cause and appreciate your understanding.

Current specifications

MODEL	DI-12-SL4	DI-12-SL02	DI-12-SL15
Accuracy	$\pm 1\%$ (500digit or less ± 5 digit)	$\pm 0.5\%$ (199 digit or less ± 1 digit)	

New specifications

MODEL	DI-12-SL4	DI-12-SL02	DI-12-SL15
Accuracy	$\pm 1.0\%$ (499digit or less, ± 5digit)	$\pm 2.0\%$ (499digit or less, ± 10digit)	$\pm 1.0\%$ (499digit or less, ± 5digit)

This is a sudden request, but we appreciate your understanding.

If you have any questions or concerns, please contact our sales department.